



TEST SIGNAL	VOUT	DIGITAL PANEL READING
14.0dBm	1.387	WAS 1.47MW SHOULD BE 1.14MW
11.0dBm	.688	

HP GENERATOR

TEST CABLE

PWR METER

ADJUST ATTEN TO GET
14.0 dBm AT END OF TEST CABLE

Peak Powr Meter, Gregorian
430 TRANSMITTER

FILE: 430pkpwr.ecw

jbh aug. 24, 2004