
## **430 TRANSMITTER SEQUENCE (7oct10)**



**Duty Cycle:**(UserReqBeam + 10)/Ipp < .06

TxIpptoBmOn=2usecs in sps software

BeamTurnOn=10 in sps softare. Hagen's box has 2 usecs.

Protect hi for 20usecs after RF in monoplexor chassis goes low.\*

Protect level high for at least 80 usecs in Hagen's box.

<sup>\*</sup> This starts after the control to platform delay